

Patient MRI Checklist

When your doctor requests an MRI scan:

- ✓ Tell your physician who requested the MRI scan that you have an implanted spinal cord stimulator device.
- ✓ Find your patient ID card. If lost, please call your Nevro HFX Care Team to request a replacement.

When making your MRI appointment:

- ✓ Provide the information below to the MRI facility when making your appointment:
 - You have a Nevro HFX spinal cord stimulator
 - If you have any other medical device implants
 - A detailed description of your device including the product number, leads, and other implant components (this information is found on your patient ID card)
 - The contact information of the physician who implanted your device
 - The Nevro HFX MRI Guidelines (available at [Nevro.com/manuals](https://www.nevro.com/manuals))
- ✓ This information will be used by your physician and radiographer to determine if you can safely undergo an MRI scan.

Preparing for your MRI appointment:

- ✓ Contact your Nevro HFX Care Team to confirm you can safely have the MRI scan scheduled.
- ✓ Before your MRI scan, your Nevro HFX Care Team will help you place your device in MRI Mode with your HFX App.
 - Please refer to the HFX App Quick Start Guide or [NevroHFX.com/en/resources/hfx-app/](https://www.NevroHFX.com/en/resources/hfx-app/) for step-by-step instructions.



Make sure your device is fully charged before your MRI scan.


Patient MRI Checklist

At your MRI appointment:

- ✓ Bring your patient ID card and HFX App to your MRI appointment.
- ✓ Your radiographer will confirm if you are in MRI Mode before your MRI scan.

After your MRI appointment:

- ✓ After your MRI scan, click “Exit MRI Mode” and follow the prompts on your HFX App to take your device out of MRI Mode and turn your stimulation back ON.
- ✓ If using Manual Mode, contact your Nevro HFX Care Team to make sure your device is programmed to your pre-MRI settings.



Do you need help?

Please contact your local HFX Care Team who will help you find a center that is compatible with spinal cord stimulators, help gather the information you need, and who can answer your questions.



MRI Guidelines for Your Radiographer

Scan the code for detailed MRI compatibility information or visit [Nevro.com/manuals](https://nevro.com/manuals).

Note: HFX and HFX iQ are approved for an MRI scan under certain conditions. Different models of implanted devices, including lead types, result in different approved scanning conditions. Please review the Nevro HFX MRI Guidelines carefully with your Nevro HFX Care Team before getting an MRI scan.

HFX, the HFX logo, HFX iQ, the HFX iQ logo, NEVRO, and the NEVRO logo are trademarks or registered trademarks of Nevro Corp.
©2024 Nevro Corp. All Rights Reserved.