

MRI Patient Checklist



The system delivering HFX™ is approved for MRI scanning under certain conditions.

When a doctor requests an MRI scan:

1. Tell the doctor who requested the MRI scan that you have an implanted spinal cord stimulation device.
2. Ask the doctor who implanted your system:
 - Can my system safely undergo the ordered MRI scan?
 - What are the specific part numbers for all implanted devices?
3. When scheduling your MRI appointment, provide the following information:
 - You have a Nevro spinal cord stimulator - It's possible that additional scanning options will be available in the future, so please refer back to the Nevro.com website to confirm the most current guidelines.
 - If you have any other medical device implants
 - A detailed description of your device including the product number, leads, and other implant components (this information is found on your patient ID card).
 - The contact information of the doctor who implanted your device
 - The MRI guidelines, which can be found at www.Nevro.com/Manuals
4. Before your MRI scan, all patients implanted with a Senza device must first conduct an impedance check. Your MRI Tech will confirm the results before your MRI.
5. Bring your patient ID card and Remote Control to the MRI appointment. If you don't have your patient ID card, please call your HFX Care Team for assistance.
6. Before your MRI scan, be sure to charge your IPG and make sure your stimulation is switched to OFF. You can turn stimulation OFF with your Remote Control.
7. After the MRI scan, turn your system back ON with your Remote Control and contact your doctor to ensure that your device is programmed to pre-MRI settings.



PTRC1000



PTRC2300



PTRC2500

Notes:

- Different models of implanted devices, including lead types, result in various approved scanning conditions.
- Please carefully review the MRI guidelines with your doctor.



Do you need help?

Please contact your local HFX Care Team who will help you find a center that is compatible with spinal cord stimulators, help gather the information you need, and who can answer your questions.



MRI Guidelines for Your Radiographer

Scan the code for detailed MRI compatibility information or visit Nevro.com/manuals.

Note: HFX is approved for an MRI scan under certain conditions. Different models of implanted devices, including lead types, result in different approved scanning conditions. Please review the Nevro HFX MRI Guidelines carefully with your Nevro HFX Care Team before getting an MRI scan.

HFX, the HFX logo, NEVRO, and the NEVRO logo are trademarks or registered trademarks of Nevro Corp. ©2024 Nevro Corp. All Rights Reserved.